



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Palen Kimball, LLC DBA PK Calibration & Validation**  
**1717 University Ave. West**  
**St. Paul, MN 55104**

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 31 October 2023

Certificate Number: AC-1814



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
AND ANSI/NCSL Z540-1-1994 (R2002)**

**Palen Kimball, LLC DBA PK Calibration & Validation**

1717 University Ave. West  
St. Paul, MN 55104

Dennis J. Brady Phone: 651-647-4501  
dbrady@palenkimball.com www.palenkimball.com

**CALIBRATION**

Valid to: **October 31, 2023**

Certificate Number: **AC-1814**

**Electrical – DC/Low Frequency**

| Parameter/Equipment                                   | Range  | Expanded Uncertainty of Measurement (+/-)  | Reference Standard, Method, and/or Equipment               |
|---|--|--|--|
| DC Voltage – Source <sup>1</sup>                      | Up to 220 mV<br>220 mV to 1 100 V  | 2.3 $\mu$ V/V + 1.6 $\mu$ V<br>2 $\mu$ V/V + 0.2 mV  | Fluke 5700A<br>Multiproduct Calibrator                     |
| DC Voltage – Measure <sup>1</sup>                     | Up to 100 mV<br>100 mV to 1 100 V  | 9 $\mu$ V/V + 0.3 $\mu$ V<br>10 $\mu$ V/V + 1 mV   | HP 3458A<br>8.5 Digit Multimeter                           |
| DC High Voltage – Measure <sup>1</sup>                | (1 to 100) kV  | 6 mV/V   | HP 3458A<br>8.5 Digit Multimeter,<br>HVI DVR-150 Divider   |
| DC Current – Source <sup>1</sup>                      | 10 $\mu$ A to 2 A  | 0.16 mA/A + 0.1 $\mu$ A  | Fluke 5700A<br>Multiproduct Calibrator                     |
|   | (2 to 11) A  | 0.66 mA/A  | Fluke 5520A<br>Multiproduct Calibrator                     |
| DC Current – Measure <sup>1</sup>                     | (0 to 1) A   | 0.14 $\mu$ A/A + 10 $\mu$ A  | HP 3458A<br>8.5 Digit Multimeter                           |
|   | (1 to 500) A   | 3.0 mA/A   | Empro HA-500-50 Shunt,<br>HP 3458A<br>8.5 Digit Multimeter |
| Resistance – Source <sup>1</sup><br>(Fixed Artifacts) | 10 $\Omega$ to 10 M $\Omega$   | 7 $\mu$ $\Omega$ / $\Omega$  | L&N<br>Standard Resistors                                  |
| Resistance – Measure <sup>1</sup>                     | Up to 2 M $\Omega$<br>(2 to 11) M $\Omega$<br>(11 to 110) M $\Omega$<br>110 M $\Omega$ to 1.1 G $\Omega$ | 0.1 m $\Omega$ / $\Omega$ + 0.24 m $\Omega$<br>0.2 m $\Omega$ / $\Omega$<br>0.8 m $\Omega$ / $\Omega$<br>8.5 m $\Omega$ / $\Omega$ | HP 3458A<br>8.5 Digit Multimeter                           |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment              | Range                | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|----------------------|---|--|
| AC Voltage – Source <sup>1</sup> | (1 to 32.999) mV     |   | Fluke 5700A<br>Multiproduct Calibrator       |
|                                  | (10 to 45) Hz        | 0.62 mV/V + 4.7 μV                        |  |
|                                  | 45 Hz to 1 kHz       | 0.12 mV/V + 4.7 μV                        |  |
|                                  | (1 to 5) kHz         | 0.12 mV/V + 4.7 μV                        |  |
|                                  | (5 to 10) kHz        | 0.12 mV/V + 4.7 μV                        |  |
|                                  | (10 to 20) kHz       | 0.15 mV/V + 4.7 μV                        |  |
|                                  | (20 to 50) kHz       | 0.78 mV/V + 4.7 μV                        |  |
|                                  | (50 to 100) kHz      | 2.7 mV/V + 9.3 μV                         |  |
|                                  | (100 to 500) kHz     | 6.2 mV/V + 39 μV                          |  |
|                                  | (33 to 329.999) mV   |   |  |
|                                  | (10 to 45) Hz        | 0.23 mV/V + 6.3 μV                        |  |
|                                  | 45 Hz to 1 kHz       | 0.11 mV/V + 6.3 μV                        |  |
|                                  | (1 to 5) kHz         | 0.11 mV/V + 6.3 μV                        |  |
|                                  | (5 to 10) kHz        | 0.11 mV/V + 6.3 μV                        |  |
|                                  | (10 to 20) kHz       | 0.12 mV/V + 6.4 μV                        |  |
|                                  | (20 to 50) kHz       | 0.27 mV/V + 6.3 μV                        |  |
|                                  | (50 to 100) kHz      | 0.62 mV/V + 25 μV                         |  |
|                                  | (100 to 500) kHz     | 1.6 mV/V + 54 μV                          |  |
|                                  | 330 mV to 3.299 99 V |   |  |
|                                  | (10 to 45) Hz        | 0.23 mV/V + 40 μV                         |  |
|                                  | 45 Hz to 1 kHz       | 0.12 mV/V + 48 μV                         |  |
|                                  | (1 to 5) kHz         | 0.12 mV/V + 47 μV                         |  |
|                                  | (5 to 10) kHz        | 0.12 mV/V + 48 μV                         |  |
|                                  | (10 to 20) kHz       | 0.15 mV/V + 48 μV                         |  |
|                                  | (20 to 50) kHz       | 0.23 mV/V + 39 μV                         |  |
|                                  | (50 to 100) kHz      | 0.54 mV/V + 98 μV                         |  |
|                                  | (100 to 500) kHz     | 1.9 mV/V + 0.47 mV                        |  |
|                                  | (3.3 to 32.999 9) V  |   |  |
|                                  | (10 to 45) Hz        | 0.23 mV/V + 0.51 mV                       |  |
|                                  | 45 Hz to 1 kHz       | 0.12 mV/V + 0.48 mV                       |  |
| (1 to 5) kHz                     | 0.12 mV/V + 0.48 mV  |   |  |
| (5 to 10) kHz                    | 0.12 mV/V + 0.48 mV  |   |  |
| (10 to 20) kHz                   | 0.19 mV/V + 0.48 mV  |   |  |
| (20 to 50) kHz                   | 0.27 mV/V + 0.47 mV  |   |  |
| (50 to 100) kHz                  | 0.7 mV/V + 1.2 mV    |   |  |
| (33 to 329.999) V                |                      |   |  |
| 45 Hz to 1 kHz                   | 0.15 mV/V + 1.9 mV   |   |  |
| (1 to 5) kHz                     | 0.15 mV/V + 4.8 mV   |   |  |
| (5 to 10) kHz                    | 0.15 mV/V + 4.8 mV   |   |  |
| (10 to 20) kHz                   | 0.19 mV/V + 4.8 mV   |   |  |
| (20 to 50) kHz                   | 0.23 mV/V + 4.7 mV   |   |  |
| (50 to 100) kHz                  | 1.6 mV/V + 39 mV     |   |  |



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

| Parameter/Equipment                    | Range  | Expanded Uncertainty of Measurement (+/-)   | Reference Standard, Method, and/or Equipment                                      |
|--|--|---|---|
| AC Voltage – Source <sup>1</sup>       | (330 to 1 020) V<br>45 Hz to 1 kHz<br>(1 to 5) kHz<br>(5 to 10) kHz  | 0.23 mV/V + 11 mV<br>0.19 mV + 9.4 mV<br>0.23 mV/V + 8.8 mV   | Fluke 5700A<br>Multiproduct Calibrator  |
| AC Voltage – Measure <sup>1</sup>      | (1 to 10) V<br>(1 to 40) Hz<br>40 Hz to 1 kHz<br>(1 to 20) kHz<br>(20 to 50) kHz<br>(50 to 100) kHz<br>(100 to 300) kHz<br>300 kHz to 1 MHz<br>(1 to 2) MHz<br>(10 to 100) V<br>(1 to 40) Hz<br>40 Hz to 1 kHz<br>(1 to 20) kHz<br>(20 to 50) kHz<br>(50 to 100) kHz<br>(100 to 300) kHz<br>300 kHz to 1 MHz<br>(100 to 700) V<br>(1 to 40) Hz<br>40 Hz to 1 kHz<br>(1 to 20) kHz<br>(20 to 50) kHz<br>(50 to 100) kHz | 0.009 5 % of reading + 46 μV<br>0.009 5 % of reading + 0.23 mV<br>0.017 % of reading + 0.23 mV<br>0.036 % of reading + 0.23 mV<br>0.093 % of reading + 0.23 mV<br>0.35 % of reading + 1.2 mV<br>1.2 % of reading + 1 mV<br>1.8 % of reading + 1 mV<br>0.024 % of reading + 4 mV<br>0.024 % of reading + 2 mV<br>0.024 % of reading + 2 mV<br>0.041 % of reading + 2 mV<br>0.14 % of reading + 2 mV<br>0.46 % of reading + 12 mV<br>1.7 % of reading + 12 mV<br>0.048 % of reading + 47 mV<br>0.048 % of reading + 23 mV<br>0.071 % of reading + 23 mV<br>0.14 % of reading + 23 mV<br>0.35 % of reading + 23 mV | HP 3458A<br>8.5 Digit Multimeter  |
| AC High Voltage – Measure <sup>1</sup> | (1 to 100) kV<br>60 Hz   | 6.6mV/V   | HP 3458A<br>8.5 Digit Multimeter,<br>HVI DVR-150 Divider                          |
| AC Current – Source <sup>1</sup>       | 9 μA to 220 mA<br>40 Hz to 1 kHz<br>10 Hz to 5 kHz<br>220 mA to 2.2 A<br>40 Hz to 1 kHz<br>(2.2 to 11) A<br>(45 to 500) Hz   | 0.33 mA/A + 0.4 μA<br>1 mA/A + 8 μA<br>0.26 mA/A + 8 μA<br>1.2 mA/A + 4 mA  | Fluke 5700A<br>Multiproduct Calibrator,<br>Fluke 5520A<br>Multiproduct Calibrator |
| AC Current – Measure <sup>1</sup>      | (0 to 1) A<br>45 Hz to 5 kHz   | 1 mA/A + 200 μA   | HP 3458A<br>8.5 Digit Multimeter  |



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

| Parameter/Equipment  | Range  | Expanded Uncertainty of Measurement (+/-)   | Reference Standard, Method, and/or Equipment   |
|--|--|---|--|
| AC Current – Measure <sup>1</sup>                            | (1 to 500) A<br>60 Hz  | 0.28 $\mu$ A/A + 400 $\mu$ A  | Empro HA-500-50 Shunt,<br>HP 3458A<br>8.5 Digit Multimeter   |
| Capacitance – Source <sup>1</sup><br>(Simulated or Artifact) | 1 kHz<br>1 nF  | 32 $\mu$ F/F  | ESI SC 1000<br>Standard Capacitor,<br>Fluke 5520A<br>Multiproduct Calibrator   |
| Capacitance – Source <sup>1</sup><br>(Artifacts)             | 190 pF to 3.3 nF<br>3.3 nF to 11 $\mu$ F<br>11 $\mu$ F to 11 mF<br>(11 to 33) mF<br>(33 to 110) mF | 0.75 mF/F + 0.01 nF<br>0.38 mF/F + 0.1 nF<br>0.69 mF/F + 10 $\mu$ F<br>1.1 mF/F + 30 $\mu$ F<br>1.7 mF/F + 0.1 mF | General Radio<br>1404A/1403D Standard<br>Capacitors, General Radio<br>1615P1 Range Extension<br>Capacitor, with General<br>Radio 1615A<br>Capacitance Bridge |
| Capacitance – Measure <sup>1</sup>                           | 1 kHz<br>1 pF to 10 $\mu$ F  | 12 $\mu$ F/F  | General Radio 1615A<br>Capacitance Bridge  |
| Inductance – Source <sup>1</sup><br>(Artifacts)              | 1 kHz<br>100 $\mu$ H to 10 H   | 81 $\mu$ H/H  | General Radio 1482B,<br>1482E, 1482H, 1482L,<br>1482P Standard Inductors,<br>with General Radio 1632A<br>Inductance Bridge                                   |
| Inductance – Measure <sup>1</sup>                            | 10 $\mu$ H to 10 H   | 0.18 mH/H   | General Radio 1632A<br>Inductance Bridge   |
| AC Power – Source <sup>1</sup>                               | 60 Hz<br>20 W to 20 kW   | 2 mW/W  | Fluke 5520A<br>Multiproduct Calibrator   |
| Oscilloscopes <sup>1</sup><br>Amplitude – Square Wave        | 45 Hz to 1 kHz<br>2.6 mVp-p to 66 Vp-p   | 4.1 mV/V + 0.15 mV  | Fluke 5520A/SC1100<br>Multiproduct Calibrator  |
| Amplitude – Leveled<br>Sine Wave<br>(50 kHz reference)       | 5 mVp-p to 5.5 Vp-p<br>50 kHz to 100 MHz<br>(100 to 300) MHz<br>(300 to 600) MHz                   | 30 mV/V + 0.5 mV<br>66 mV/V + 0.5 mV<br>61 mV/V + 0.5 mV  |  |
| Flatness<br>(50 kHz reference)                               | 50 kHz to 100 MHz<br>(100 to 300) MHz<br>(300 to 600) MHz<br>600 MHz to 1.1 GHz                    | 3 mV/V + 0.5 mV<br>6 mV/V + 0.5 mV<br>66 mV/V + 0.5 mV<br>61 mV/V + 0.5 mV  |  |



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

| Parameter/Equipment  | Range  | Expanded Uncertainty of Measurement (+/-)      | Reference Standard, Method, and/or Equipment  |
|--|--|--|---|
| Oscilloscopes <sup>1</sup><br>Time Marker<br>into 50 Ω load                          | 2 ns to 1 μs<br>2 μs to 50 ms<br>100 ms to 5 s   | 4.4 μs/s<br>5.1 μs/s + 15 ms<br>15 μs/s + 1 ms | Fluke 5520A/SC1100<br>Multiproduct Calibrator |
| Rise Time  | < 2 ns   | 450 ps   |   |
| Electrical Simulation of<br>Thermocouple Indicators –<br>Source/Measure <sup>1</sup> | Type J<br>(-210 to 1 200) °C<br>Type K<br>(-200 to 1 372) °C<br>Type T<br>(-250 to 400) °C<br>Type E<br>(-200 to 1 000) °C                             | 0.27 °C<br>0.45 °C<br>0.72 °C<br>0.45 °C       | Fluke 5520A<br>Multiproduct Calibrator        |
| Electrical Simulation of RTD<br>Indicators – Source <sup>1</sup>                     | Pt 385 (100 Ω)<br>(-200 to 630) °C<br>Pt 385 (200 Ω)<br>(-200 to 630) °C<br>Pt 3926 (100 Ω)<br>(-200 to 630) °C<br>Pt 3916 (100 Ω)<br>(-200 to 630) °C | 0.18 °C<br>0.24 °C<br>0.18 °C<br>0.37 °C       | Fluke 5520A<br>Multiproduct Calibrator        |

**Length – Dimensional Metrology**

| Parameter/Equipment           | Range       | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------|-------------|---|--|
| Calipers <sup>1,3</sup>       | Up to 24 in | 11 μin/in                                 | Gage Blocks                                  |
| Micrometers <sup>1,3</sup>    | Up to 24 in | 11 μin/in                                 |  |
| Length Standards <sup>2</sup> | Up to 24 in | (12 + 4.1L) μin                           | Mahr ULM 600E<br>Length Measuring Machine    |

### Mass and Mass Related

| Parameter/Equipment                  | Range  | Expanded Uncertainty of Measurement (+/-)   | Reference Standard, Method, and/or Equipment |
|--------------------------------------|--|---|--|
| Pressure/Vacuum Gages <sup>1</sup>   | (-5 to 5) psi<br>(-15 to 30) psi<br>Up to 30 inHg<br>Up to 10 inH <sub>2</sub> O<br>Up to 100 psi<br>Up to 300 psi<br>Up to 500 psi<br>Up to 7 500 psi<br>Up to 10 000 psi | 0.003 psi<br>0.04 psi<br>0.04 inHg<br>0.01 inH <sub>2</sub> O<br>0.054 psi<br>0.2 psi<br>0.41 psi<br>5 psi<br>8.4 psi | Pressure Calibrators with Pressure Modules   |
| Absolute Pressure Gages <sup>1</sup> | Up to 100 psia   | 0.07 psi  |  |

### Thermodynamic

| Parameter/Equipment                | Range            | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment          |
|------------------------------------|------------------|---|---|
| Temperature – Measure <sup>1</sup> | (-200 to 420) °C | 0.02 °C                                   | Azonix AI101 Thermometer, Burns Engineering 12001 PRT |
| Humidity – Measure <sup>1</sup>    | Up to 90 %RH     | 1.4 % RH                                  | Vaisala HM70/HMP77B Digital Indicator/Probe           |

### Time and Frequency

| Parameter/Equipment              | Range            | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment          |
|----------------------------------|------------------|---|---|
| Frequency – Source <sup>1</sup>  | 1 MHz<br>10 MHz  | 65 pHz/Hz<br>65 pHz/Hz                    | WWVB Phase Chart Recorder, GPS Disciplined Oscillator |
| Frequency – Measure <sup>1</sup> | 1 mHz to 350 MHz | 1.8 μHz/Hz                                | Agilent 53230A Universal Frequency Counter            |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = length in inches,  $R$  = resolution of unit under test.
3. The Best Measurement Capability presented here does not include the Resolution contributor. The value of  $0.6R$  will be added to the Measurement Uncertainty at the time of calibration.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1814.



R. Douglas Leonard Jr., VP, PILR SBU

