



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Palen Kimball, LLC DBA PK Calibration & Validation
1717 University Ave. West
St. Paul, MN 55104

Fulfills the requirements of

ISO/IEC 17025:2017

and

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 31 October 2025

Certificate Number: AC-1814



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Palen Kimball, LLC DBA PK Calibration & Validation

1717 University Ave. West

St. Paul, MN 55104

Randy Hiber Phone: 612-369-2375 rhiber@palenkimball.com

CALIBRATION

Valid to: **October 31, 2025**

Certificate Number: **AC-1814**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ¹	Up to 220 mV 220 mV to 1 100 V	2.3 μ V/V + 1.6 μ V 2 μ V/V + 0.2 mV	Multiproduct Calibrator
DC Voltage – Measure ¹	Up to 100 mV 100 mV to 1 100 V	9 μ V/V + 0.3 μ V 10 μ V/V + 1 mV	8.5 Digit Multimeter
DC High Voltage – Measure ¹	(1 to 100) kV	6 mV/V	8.5 Digit Multimeter, High Voltage Divider
DC Current – Source ¹	10 μ A to 2 A	0.16 mA/A + 0.1 μ A	Multiproduct Calibrator
DC Current – Source ¹	(2 to 11) A	0.66 mA/A	Multiproduct Calibrator
DC Current – Measure ¹	(0 to 1) A	0.14 μ A/A + 10 μ A	8.5 Digit Multimeter
Resistance – Source ¹ (Fixed Artifacts)	10 Ω to 10 M Ω	7 μ Ω / Ω	Standard Resistors
Resistance – Measure ¹	Up to 2 M Ω (2 to 11) M Ω (11 to 110) M Ω 110 M Ω to 1.1 G Ω	0.1 m Ω / Ω + 0.24 m Ω 0.2 m Ω / Ω 0.8 m Ω / Ω 8.5 m Ω / Ω	8.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(1 to 32.999) mV		Multiproduct Calibrator
	(10 to 45) Hz	0.62 mV/V + 4.7 μV	
	45 Hz to 1 kHz	0.12 mV/V + 4.7 μV	
	(1 to 5) kHz	0.12 mV/V + 4.7 μV	
	(5 to 10) kHz	0.12 mV/V + 4.7 μV	
	(10 to 20) kHz	0.15 mV/V + 4.7 μV	
	(20 to 50) kHz	0.78 mV/V + 4.7 μV	
	(50 to 100) kHz	2.7 mV/V + 9.3 μV	
	(100 to 500) kHz	6.2 mV/V + 39 μV	
	(33 to 329.999) mV		
	(10 to 45) Hz	0.23 mV/V + 6.3 μV	
	45 Hz to 1 kHz	0.11 mV/V + 6.3 μV	
	(1 to 5) kHz	0.11 mV/V + 6.3 μV	
	(5 to 10) kHz	0.11 mV/V + 6.3 μV	
	(10 to 20) kHz	0.12 mV/V + 6.4 μV	
	(20 to 50) kHz	0.27 mV/V + 6.3 μV	
	(50 to 100) kHz	0.62 mV/V + 25 μV	
	(100 to 500) kHz	1.6 mV/V + 54 μV	
	330 mV to 3.299 99 V		
	(10 to 45) Hz	0.23 mV/V + 40 μV	
	45 Hz to 1 kHz	0.12 mV/V + 48 μV	
	(1 to 5) kHz	0.12 mV/V + 47 μV	
	(5 to 10) kHz	0.12 mV/V + 48 μV	
	(10 to 20) kHz	0.15 mV/V + 48 μV	
	(20 to 50) kHz	0.23 mV/V + 39 μV	
	(50 to 100) kHz	0.54 mV/V + 98 μV	
	(100 to 500) kHz	1.9 mV/V + 0.47 mV	
	(3.3 to 32.999 9) V		
	(10 to 45) Hz	0.23 mV/V + 0.51 mV	
	45 Hz to 1 kHz	0.12 mV/V + 0.48 mV	
(1 to 5) kHz	0.12 mV/V + 0.48 mV		
(5 to 10) kHz	0.12 mV/V + 0.48 mV		
(10 to 20) kHz	0.19 mV/V + 0.48 mV		
(20 to 50) kHz	0.27 mV/V + 0.47 mV		
(50 to 100) kHz	0.7 mV/V + 1.2 mV		
(33 to 329.999) V			
45 Hz to 1 kHz	0.15 mV/V + 1.9 mV		
(1 to 5) kHz	0.15 mV/V + 4.8 mV		
(5 to 10) kHz	0.15 mV/V + 4.8 mV		
(10 to 20) kHz	0.19 mV/V + 4.8 mV		
(20 to 50) kHz	0.23 mV/V + 4.7 mV		
(50 to 100) kHz	1.6 mV/V + 39 mV		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(330 to 1 020) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.23 mV/V + 11 mV 0.19 mV + 9.4 mV 0.23 mV/V + 8.8 mV	Multiproduct Calibrator
AC Voltage – Measure ¹	(1 to 10) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.009 5 % of reading + 46 μV 0.009 5 % of reading + 0.23 mV 0.017 % of reading + 0.23 mV 0.036 % of reading + 0.23 mV 0.093 % of reading + 0.23 mV 0.35 % of reading + 1.2 mV 1.2 % of reading + 1 mV 1.8 % of reading + 1 mV 0.024 % of reading + 4 mV 0.024 % of reading + 2 mV 0.024 % of reading + 2 mV 0.041 % of reading + 2 mV 0.14 % of reading + 2 mV 0.46 % of reading + 12 mV 1.7 % of reading + 12 mV 0.048 % of reading + 47 mV 0.048 % of reading + 23 mV 0.071 % of reading + 23 mV 0.14 % of reading + 23 mV 0.35 % of reading + 23 mV	8.5 Digit Multimeter
AC High Voltage – Measure ¹	(1 to 100) kV 60 Hz	6.6 mV/V	8.5 Digit Multimeter, High Voltage Divider
AC Current – Source ¹	9 μA to 220 mA 40 Hz to 1 kHz 10 Hz to 5 kHz 220 mA to 2.2 A 40 Hz to 1 kHz (2.2 to 11) A (45 to 500) Hz	0.33 mA/A + 0.4 μA 1 mA/A + 8 μA 0.26 mA/A + 8 μA 1.2 mA/A + 4 mA	Multiproduct Calibrator
AC Current – Measure ¹	Up to 1 A 45 Hz to 5 kHz	1 mA/A + 0.2 mA	8.5 Digit Multimeter
Capacitance – Source ¹ (Simulated or Artifact)	1 kHz 1 nF	32 μF/F	Standard Capacitor or Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ¹ (Artifacts)	190 pF to 3.3 nF 3.3 nF to 11 μF 11 μF to 11 mF (11 to 33) mF (33 to 110) mF	0.75 mF/F + 10 pF 0.38 mF/F + 0.1 nF 0.69 mF/F + 10 μF 1.1 mF/F + 30 μF 1.7 mF/F + 0.1 mF	Standard Capacitors, Decade Capacitor
Inductance – Source ¹ (Artifacts)	1 kHz 100 μH to 10 H	81 μH/H	Standard Inductors
AC Power – Source ¹	60 Hz 20 W to 20 kW	2 mW/W	Multiproduct Calibrator
Oscilloscopes ¹ Amplitude – Square Wave	45 Hz to 1 kHz 2.6 mVp-p to 66 Vp-p	4.1 mV/V + 0.15 mV	Multiproduct Calibrator with 1.1 GHz Scope Option
Amplitude – Leveled Sine Wave (50 kHz reference)	5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	30 mV/V + 0.5 mV 66 mV/V + 0.5 mV 61 mV/V + 0.5 mV	
Flatness (50 kHz reference)	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 600 MHz to 1.1 GHz	3 mV/V + 0.5 mV 6 mV/V + 0.5 mV 66 mV/V + 0.5 mV 61 mV/V + 0.5 mV	
Time Marker into 50 Ω load	2 ns to 1 μs 2 μs to 50 ms 100 ms to 5 s	4.4 μs/s 5.1 μs/s + 15 ms 15 μs/s + 1 ms	
Rise Time	< 2 ns	450 ps	
Electrical Simulation of Thermocouple Indicators – Source/Measure ¹	Type J (-210 to 1 200) °C Type K (-200 to 1 372) °C Type T (-250 to 400) °C Type E (-200 to 1 000) °C	0.27 °C 0.45 °C 0.72 °C 0.45 °C	Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicators – Source ¹	Pt 385 (100 Ω) (-200 to 630) °C	0.18 °C	Multiproduct Calibrator
	Pt 385 (200 Ω) (-200 to 630) °C	0.24 °C	
	Pt 3926 (100 Ω) (-200 to 630) °C	0.18 °C	
	Pt 3916 (100 Ω) (-200 to 630) °C	0.37 °C	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers ^{1,2}	Up to 24 in	(570 + 14L) μin	Gage Blocks
Micrometers ^{1,2}	Up to 24 in	(59 + 5.3L) μin	Gage Blocks
Length Standards ^{1,2}	Up to 24 in	(8 + 4.1L) μin	Universal Length Measuring Machine, Gage Blocks
Dial/Digital Indicators ¹	Up to 1 in	600 μin	Indicator Calibrator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Gages ¹	Up to 10 inH ₂ O	0.07 inH ₂ O	Heise Pressure Indicator with Pressure Modules
	Up to 100 psig	0.16 psi	
	Up to 300 psig	0.2 psi	
	Up to 1 000 psig	0.41 psi	
	Up to 7 500 psig	5 psi	
	Up to 10 000 psig	7 psi	
Compound Gages ¹ (Pneumatic)	(-15 to 300) psig	0.15 psi	Druck Pressure Indicator with Pressure Modules
Absolute Pressure Gages ¹ (Pneumatic)	Up to 100 psia	0.07 psi	Heise Pressure Indicator with Absolute Pressure Module

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure ¹	(-200 to 420) °C	0.02 °C	Thermometer Readout, PRT
Humidity – Measure ¹	(11 to 80) %RH (80 to 90) %RH	1.4 % RH 2.1 %RH	Digital Indicator with Humidity Probe


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source ¹	1 MHz 10 MHz	3.7 pHz/Hz 3.7 pHz/Hz	GPS Disciplined Oscillator
Frequency – Measure ¹	1 mHz to 350 MHz	0.002 % of reading	Universal Frequency Counter, GPS Disciplined Oscillator
Stopwatches, Timers ¹	Up to 86 400 s	37 ms/d	Timometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = length in inches.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1814.



Jason Stine, Vice President